


ovesco training campus
TÜBINGEN
CARY
RENNES
CHENNAI


OTSC®neo System – Technical specifications

1. Maximum compatible endoscope diameter
2. OTSC®neo cap depth (3 mm or 6 mm)
3. Teeth configuration (a, t or gc)

12 / 6 t




1 OTSC®neo cap options: 4 sizes



	mini	11	12	14
Compatible endoscope OD [mm] (a)	8.5 – 10	8.5 – 11	10.5 – 12	11.5 – 14
Endoscope type*	Diagnostic EGD Scope	Diagnostic/ 1T EGD Scope	1T/2T EGD Scope; Pediatric Colonoscope	Adult & pediatric Colonoscopes
Maximum cap OD [mm] (b)	14.6	16	17.4	20.1
Thread length [cm]	165	165	165 or 220	220

2 OTSC®neo cap depth: 2 sizes


3 mm depth



Less tissue capture, enhanced visibility and maneuverability

Available in 11, 12 & 14 size caps

6 mm depth



Standard depth for maximum tissue capture

Available in mini, 11, 12 & 14 size caps

3 OTSC®neo teeth configuration: 3 types

Type a



Blunt teeth

Mainly used for upper/lower GI bleeding

Type t



Teeth with small spikes

Mainly used for upper/lower GI bleeding, perforation & fistula closure

Type gc



Teeth with long spikes

Mainly used for gastric perforation & fistula closure *(only in size 12 cap)*

Tissue management devices

OTSC® Anchor



For mobilization of tissue, such as fibrotic or hard tissue, e.g. fistulas or chronic ulcers

OTSC® TwinGrasper®



For approximation of opposing edges of a lesion, e.g. perforations

* This information is provided as general guidance for endoscope compatibility. OD may vary, based on scope manufacturing tolerance and repair history. Confirm the actual OD of your institution's endoscopes prior to using the OTSC®neo System to ensure compatibility (e.g. by using the ScopeCheck).

More information on OTSC®neo

OTSC®neo infosheet-technical-specifications_EN Rev01_2026-08-12

Ovesco Endoscopy AG | Friedrich-Miescher-Str. 9 | 72076 Tübingen | Germany | +49 (0)7071 96528-160 | service@ovesco.com | www.ovesco.com